

TEST CERTIFICATES

We offer test reports or test certificates acc. to DIN EN 10204.

- | | |
|--|---------------|
| 1. Declaration of compliance with the order acc. to DIN EN 10204-2.1 | 45,- € |
| Manufacturer's declaration of compliance with the order without test results. | |
| 2. Test report acc. to DIN EN 10204-2.2 (batch certificate) | 80,- € |
| Manufacturer's declaration of compliance with the order, with test results based on non specific inspection. | |
| 3. Inspection certificate acc. to DIN EN 10204-3.1 | 80,- € |
| Manufacturer's declaration of compliance with the order, with test results based on specific inspection. | |

The test unit and the execution of the test are determined in the product specification, in official or technical prescriptions and/ or order. The certificate is confirmed by a person independent of production and named by the manufacturer.

List of individual tests per measuring point	25,- €
calibration in „Kryostat“ bath: temperature range -50°C up to +50°C	
calibration in oil bath: temperature range +60°C up to +200°C	
calibration in „Trockenblock-Kalibrator“: temperature range -30°C up to +165°C, +100°C up to +1100°C	
response time in water: determination of 0,1-time, 0,5-time and 0,9-time	
response time in air: determination of 0,1-time, 0,5-time and 0,9-time	